

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
 Data File : VU055675.D
 Acq On : 05 Oct 2023 16:18
 Operator : MD/SY
 Sample : 04744-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH580

Quant Time: Oct 06 07:10:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 06 06:49:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	10944	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	16861	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5934	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	191	0.029	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	6.000%#
4) 1,2-Dichloroethane-d4	5.709	65	4450	0.542	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.000%
7) 1,2-Dichloropropane-d6	6.687	67	5340	0.539	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	108.000%
8) Toluene-d8	7.887	98	5925	0.265	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	52.000%#
10) 1,1,2,2-Tetrachloroeth...	10.745	84	4578	0.579	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	116.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100523\
Data File : VU055675.D
Acq On : 05 Oct 2023 16:18
Operator : MD/SY
Sample : 04744-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH580

Quant Time: Oct 06 07:10:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100523.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 06 06:49:44 2023
Response via : Initial Calibration

