

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\
 Data File : VU027464.D
 Acq On : 05 Oct 2018 19:33
 Operator : MD/SY
 Sample : J5165-09DL 20X
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EP-MW-H-092718DL

Quant Time: Oct 06 01:36:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	112477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	188320	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	183574	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	105946	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	105213	56.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.04%	
35) Dibromofluoromethane	4.89	113	74820	57.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.94%	
50) Toluene-d8	7.57	98	291571	60.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.72%	
62) 4-Bromofluorobenzene	10.31	95	109207	60.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.66%	
Target Compounds						
16) Acetone	2.32	43	4653	4.194	ug/l	97
73) Isopropylbenzene	10.17	105	9697	1.260	ug/l	97
78) n-propylbenzene	10.59	91	22200	2.401	ug/l	100
80) 1,3,5-Trimethylbenzene	10.77	105	9419	1.504	ug/l	98
84) 1,2,4-Trimethylbenzene	11.15	105	227971	35.624	ug/l	100
85) sec-Butylbenzene	11.32	105	11471	1.503	ug/l	100
86) p-Isopropyltoluene	11.48	119	11448	1.752	ug/l	97
89) n-Butylbenzene	11.89	91	8588	3.140	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\
Data File : VU027464.D
Acq On : 05 Oct 2018 19:33
Operator : MD/SY
Sample : J5165-09DL 20X
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EP-MW-H-092718DL

Quant Time: Oct 06 01:36:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

