

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\
 Data File : VU027466.D
 Acq On : 05 Oct 2018 20:20
 Operator : MD/SY
 Sample : J5237-10
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FBSOIL

Quant Time: Oct 06 01:58:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	109784	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	184179	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	178125	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	97028	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	102981	56.18	ug/l	0.00
Spiked Amount			Recovery	=	112.36%	
35) Dibromofluoromethane	4.89	113	73474	58.21	ug/l	0.00
Spiked Amount			Recovery	=	116.42%	
50) Toluene-d8	7.57	98	279983	59.27	ug/l	0.00
Spiked Amount			Recovery	=	118.54%	
62) 4-Bromofluorobenzene	10.31	95	104688	59.63	ug/l	0.00
Spiked Amount			Recovery	=	119.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\
Data File : VU027466.D
Acq On : 05 Oct 2018 20:20
Operator : MD/SY
Sample : J5237-10
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FBSOIL

Quant Time: Oct 06 01:58:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

