

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\  
 Data File : VU027467.D  
 Acq On : 05 Oct 2018 20:43  
 Operator : MD/SY  
 Sample : J5268-10  
 Misc : 5.0mL/MSVOA U/WATER  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 TRIP-BLANK

Quant Time: Oct 06 02:00:44 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\82U100118W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Oct 02 02:04:34 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 4.99  | 168  | 112505   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.89  | 114  | 191044   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.09  | 117  | 184293   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 11.48 | 152  | 99431    | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 33) 1,2-Dichloroethane-d4   | 5.31  | 65   | 105996   | 56.42 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 112.84% |          |
| 35) Dibromofluoromethane    | 4.89  | 113  | 76068    | 58.10 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 116.20% |          |
| 50) Toluene-d8              | 7.57  | 98   | 291974   | 59.58 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 119.16% |          |
| 62) 4-Bromofluorobenzene    | 10.31 | 95   | 105667   | 58.02 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 116.04% |          |

| Target Compounds | Qvalue |
|------------------|--------|
|------------------|--------|

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\  
Data File : VU027467.D  
Acq On : 05 Oct 2018 20:43  
Operator : MD/SY  
Sample : J5268-10  
Misc : 5.0mL/MSVOA U/WATER  
ALS Vial : 31 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
TRIP-BLANK

Quant Time: Oct 06 02:00:44 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\82U100118W.M  
Quant Title : SW846 8260  
QLast Update : Tue Oct 02 02:04:34 2018  
Response via : Initial Calibration

