

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\
 Data File : VU027477.D
 Acq On : 06 Oct 2018 00:35
 Operator : MD/SY
 Sample : J5205-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TR-05-100418

Quant Time: Oct 06 03:57:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	107056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	185341	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	180528	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	97682	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	103176	57.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.44%	
35) Dibromofluoromethane	4.89	113	73123	57.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.14%	
50) Toluene-d8	7.57	98	285133	59.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.96%	
62) 4-Bromofluorobenzene	10.31	95	104806	59.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.64%	
Target Compounds						
16) Acetone	2.32	43	8442	7.994	ug/l	99
18) Methyl Acetate	2.62	43	2821	1.171	ug/l	95
20) Methylene Chloride	2.70	84	28092	19.096	ug/l	97
43) Isopropyl Acetate	5.55	43	115137	28.504	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\
Data File : VU027477.D
Acq On : 06 Oct 2018 00:35
Operator : MD/SY
Sample : J5205-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TR-05-100418

Quant Time: Oct 06 03:57:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

