

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\
 Data File : VU027478.D
 Acq On : 06 Oct 2018 00:59
 Operator : MD/SY
 Sample : J5174-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 T2-GRAVEL

Quant Time: Oct 06 04:00:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	105532	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	183431	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	175912	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	97165	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	102714	58.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.58%	
35) Dibromofluoromethane	4.89	113	72338	57.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.08%	
50) Toluene-d8	7.57	98	279523	59.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.82%	
62) 4-Bromofluorobenzene	10.31	95	100862	57.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.36%	
Target Compounds						
16) Acetone	2.32	43	14908	14.320	ug/l	98
20) Methylene Chloride	2.70	84	27620	19.047	ug/l	98
43) Isopropyl Acetate	5.55	43	101088	25.287	ug/l	98
95) Naphthalene	13.75	128	19797	4.265	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\
Data File : VU027478.D
Acq On : 06 Oct 2018 00:59
Operator : MD/SY
Sample : J5174-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
T2-GRAVEL

Quant Time: Oct 06 04:00:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

