

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100618\
 Data File : VU027482.D
 Acq On : 06 Oct 2018 02:32
 Operator : MD/SY
 Sample : J5262-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 N48964-BS

Quant Time: Oct 06 04:17:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	111347	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	190537	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	181515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.49	152	98590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	104221	56.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.10%	
35) Dibromofluoromethane	4.89	113	73818	56.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.06%	
50) Toluene-d8	7.57	98	289766	59.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.58%	
62) 4-Bromofluorobenzene	10.31	95	105834	58.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.54%	
Target Compounds						
16) Acetone	2.32	43	4977	4.531	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100618\
Data File : VU027482.D
Acq On : 06 Oct 2018 02:32
Operator : MD/SY
Sample : J5262-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
N48964-BS

Quant Time: Oct 06 04:17:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

