

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100719\
 Data File : VU034959.D
 Acq On : 07 Oct 2019 14:58
 Operator : JC/SP
 Sample : K5196-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 09 04:18:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 09 03:56:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	196071	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	191096	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.46	152	81042	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	52123	4.50	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.00%
7) Chloroethane-d5	1.67	69	47656	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.00%
11) 1,1-Dichloroethene-d2	2.26	63	78350	3.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	64.20%
20) 2-Butanone-d5	4.18	46	158871	46.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.10%
24) Chloroform-d	4.62	84	94412	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
26) 1,2-Dichloroethane-d4	5.29	65	59459	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	5.31	84	215535	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	6.31	67	73306	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.54	98	198393	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
43) trans-1,3-Dichloropropene-	7.84	79	22838	4.24	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.80%
46) 2-Hexanone-d5	8.30	63	121637	47.44	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.88%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	60249	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	11.84	152	72496	5.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.00%


Target Compounds				Ovalue
25) Chloroform	4.65	83	16766	0.557 ug/L 95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100719\
Data File : VU034959.D
Acq On : 07 Oct 2019 14:58
Operator : JC/SP
Sample : K5196-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 09 04:18:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 09 03:56:49 2019
Response via : Initial Calibration





#25
 Chloroform
 Concen: 0.557 ug/L
 RT: 4.65 min Scan# 1108
 Delta R.T. 0.00 min
 Lab File: VU034959.D
 Acq: 07 Oct 2019 14:58

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Tot Ion: 83 Resp: 16766
 Ion Ratio Lower Upper
 83 100
 85 69.0 45.4 84.4

