

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100721\
 Data File : VU045223.D
 Acq On : 07 Oct 2021 21:28
 Operator : SY/MD
 Sample : M4023-20
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AK9

Quant Time: Oct 07 22:20:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 07 22:18:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	287188	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.423	117	277857	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.815	152	121394	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	129485	41.186	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	82.380%	
7) Chloroethane-d5	1.893	69	103258	41.447	ug/L	-0.02
Spiked Amount	50.000	Range 70 - 130	Recovery	=	82.900%	
11) 1,1-Dichloroethene-d2	2.562	63	163399	30.949	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.900%	
21) 2-Butanone-d5	4.626	46	188831	75.974	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	75.970%	
24) Chloroform-d	5.067	84	201791	40.639	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	81.280%	
26) 1,2-Dichloroethane-d4	5.706	65	144352	42.919	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.840%	
32) Benzene-d6	5.732	84	425321	44.113	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.220%	
36) 1,2-Dichloropropane-d6	6.697	67	136658	43.469	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	86.940%	
41) Toluene-d8	7.902	98	370602	44.541	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.080%	
43) trans-1,3-Dichloroprop...	8.182	79	54110	38.693	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	77.380%	
47) 2-Hexanone-d5	8.639	63	134514	77.210	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	77.210%	
56) 1,1,2,2-Tetrachloroeth...	10.761	84	202625	41.390	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	82.780%	
66) 1,2-Dichlorobenzene-d4	12.198	152	128954	48.812	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	97.620%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.571	96	1903	0.856	ug/L	# 1
34) Trichloroethene	6.546	95	21182	8.557	ug/L	99
42) Toluene	7.973	91	5853	0.559	ug/L	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100721\
Data File : VU045223.D
Acq On : 07 Oct 2021 21:28
Operator : SY/MD
Sample : M4023-20
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AK9

Quant Time: Oct 07 22:20:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 07 22:18:46 2021
Response via : Initial Calibration





#12

1,1-Dichloroethene

Concen: 0.856 ug/L

RT: 2.571 min Scan# 383

Delta R.T. -0.010 min





Lab File: VU045223.D

Acq: 07 Oct 2021 21:28

Tgt Ion: 96 Resp: 1903

Ion	Ratio	Lower	Upper
96	100		
61	905.4	131.3	243.9#
63	6907.6	102.5	190.3#

Instrument :
MSVOA_U
ClientSampled :
C0AK9



#34

Trichloroethene

Concen: 8.557 ug/L

RT: 6.546 min Scan# 1619




Delta R.T. -0.000 min


Lab File: VU045223.D

Acq: 07 Oct 2021 21:28

Tgt Ion: 95 Resp: 21182

Ion	Ratio	Lower	Upper
95	100		
97	63.9	44.4	82.4
132	91.0	62.9	116.9
130	93.8	65.0	120.8





#42
Toluene
Concen: 0.559 ug/L
RT: 7.973 min Scan# 2063
Delta R.T. -0.000 min
Lab File: VU045223.D
Acq: 07 Oct 2021 21:28

Instrument :
MSVOA_U
ClientSampleId :
C0AK9

Tgt Ion: 91 Resp: 5853
Ion Ratio Lower Upper
91 100
92 55.5 40.5 75.1

