



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100721\
Data File : VU045197.D
Acq On : 07 Oct 2021 09:01
Operator : SY/MD
Sample : BFB008
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB008

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Title : VOC Analysis
Last Update : Mon Oct 11 01:47:08 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2878

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.5	10999	PASS
75	95	30	80	53.9	28936	PASS
95	95	100	100	100.0	53675	PASS
96	95	5	9	6.8	3624	PASS
173	174	0.00	2	0.4	148	PASS
174	95	50	120	72.4	38872	PASS
175	174	5	9	7.6	2970	PASS
176	174	95	101	95.0	36947	PASS
177	176	5	9	6.6	2426	PASS

m/z	Abundance
43.80	265.0

Instrument :
MSVOA_U
ClientSampleId :
BFB008

m/z	Abundance
36.00	551.0
37.00	2831.0
37.90	2516.0
38.90	1153.0
39.80	112.0
43.90	605.0
44.90	418.0
47.00	683.0
47.80	342.0
49.00	2695.0
49.90	10753.0
50.90	3513.0
51.90	200.0
54.90	194.0
55.90	970.0
56.90	1682.0
60.00	480.0
60.90	2874.0
61.90	3110.0
62.90	2079.0
63.90	182.0
64.90	161.0
66.90	191.0
67.90	5492.0
68.90	5710.0
69.90	528.0
71.90	322.0
72.90	2517.0
73.90	9503.0
75.00	27792.0
75.90	2519.0
76.90	323.0
77.90	234.0
78.90	1636.0
79.90	560.0
80.80	1918.0
81.80	385.0
86.90	2138.0
87.80	1948.0
90.90	229.0
91.90	1498.0
92.90	2281.0
94.00	5791.0
95.00	50464.0
96.00	3400.0
96.80	144.0
103.80	235.0
104.70	100.0
105.80	258.0
115.80	216.0
116.80	304.0
117.80	182.0
118.90	225.0
127.80	125.0
129.80	167.0
140.80	394.0
142.80	418.0
171.80	220.0
173.00	443.0
173.90	36336.0
174.90	2827.0
175.90	34288.0
176.90	2356.0

Instrument :
MSVOA_U
ClientSampleId :
BFB008

m/z	Abundance
36.00	582.0
37.00	3105.0
38.00	2649.0
38.90	1175.0
39.80	105.0
43.80	604.0
44.90	466.0
47.00	722.0
47.90	372.0
48.90	2875.0
49.90	11268.0
50.90	3939.0
51.90	208.0
54.90	224.0
55.90	1040.0
56.90	1818.0
60.00	491.0
60.90	2965.0
61.90	3251.0
62.90	2157.0
63.90	180.0
64.90	183.0
66.90	200.0
67.90	5957.0
68.90	6090.0
69.90	568.0
71.90	344.0
72.90	2694.0
73.90	10548.0
75.00	29704.0
75.90	2624.0
76.90	341.0
78.00	275.0
78.90	1788.0
79.90	574.0
80.80	1983.0
81.80	415.0
86.90	2332.0
87.90	2120.0
90.90	216.0
91.90	1631.0
92.90	2520.0
94.00	6387.0
95.00	55328.0
96.00	3732.0
96.90	115.0
103.80	214.0
104.70	104.0
105.80	296.0
115.90	209.0
116.80	324.0
117.80	189.0
118.90	257.0
127.80	146.0
129.80	190.0
140.80	444.0
142.80	441.0
147.70	118.0
171.80	241.0
173.90	40072.0
174.90	3073.0
175.90	38048.0
176.90	2495.0

Instrument :
MSVOA_U
ClientSampleId :
BFB008

m/z	Abundance
35.90	559.0
37.00	3003.0
38.00	2516.0
38.90	1108.0
39.80	103.0
43.80	609.0
44.90	471.0
47.00	677.0
47.90	386.0
48.90	2782.0
50.00	10977.0
50.90	3797.0
51.90	196.0
54.90	202.0
55.90	980.0
56.90	1715.0
59.90	467.0
60.90	2831.0
62.00	3020.0
63.00	2167.0
63.80	206.0
64.80	183.0
67.00	203.0
67.90	6006.0
68.90	5967.0
69.90	516.0
71.90	332.0
72.90	2641.0
73.90	10528.0
75.00	29312.0
76.00	2465.0
76.90	322.0
78.00	249.0
78.80	1780.0
79.90	540.0
80.90	1809.0
81.80	377.0
86.90	2330.0
87.90	2251.0
90.90	193.0
91.90	1627.0
92.90	2515.0
94.00	6548.0
95.00	55232.0
96.00	3741.0
103.80	193.0
104.80	103.0
105.80	279.0
115.90	191.0
116.80	307.0
117.80	195.0
118.80	280.0
127.70	157.0
129.80	176.0
134.80	108.0
140.80	425.0
142.80	436.0
147.70	122.0
171.80	235.0
173.90	40208.0
174.90	3010.0
175.90	38504.0
176.90	2426.0

Instrument :
MSVOA_U
ClientSampleId :
BFB008