

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100818\
 Data File : VU027513.D
 Acq On : 08 Oct 2018 20:44
 Operator : MD/SY
 Sample : J5298-13
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E62W6

Quant Time: Oct 09 09:26:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100518WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 09 05:21:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.89	114	179260	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.09	117	163808	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.49	152	70152	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	54589	38.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.40%
7) Chloroethane-d5	1.68	69	47223	43.45	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.90%
11) 1,1-Dichloroethene-d2	2.27	63	79228	29.94	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.88%#
21) 2-Butanone-d5	4.18	46	134069	117.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.72%
24) Chloroform-d	4.65	84	109270	48.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.86%
26) 1,2-Dichloroethane-d4	5.31	65	76882	48.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.24%
32) Benzene-d6	5.34	84	212592	45.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.56%
36) 1,2-Dichloropropane-d6	6.33	67	80228	49.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.78%
41) Toluene-d8	7.57	98	175382	41.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.58%
43) trans-1,3-Dichloropropene-	7.85	79	33904	43.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.50%
47) 2-Hexanone-d5	8.31	63	92706	114.15	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	114.15%
57) 1,1,2,2-Tetrachloroethane-	10.43	84	110058	52.34	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.68%
64) 1,2-Dichlorobenzene-d4	11.86	152	67745	49.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.72%

Target Compounds

					Ovalue
42) Toluene	7.64	91	83379	15.437 ug/L	99
53) m,p-Xylene	9.38	106	1748	0.796 ug/L	95
54) o-xylene	9.78	106	1384	0.654 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100818\
Data File : VU027513.D
Acq On : 08 Oct 2018 20:44
Operator : MD/SY
Sample : J5298-13
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E62W6

Quant Time: Oct 09 09:26:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100518WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 09 05:21:41 2018
Response via : Initial Calibration

