




Data File : VU034983.D
Acq On : 08 Oct 2019 11:02
Operator : JC/SP
Sample : MDL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Oct 09 07:39:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 09 07:15:09 2019
Response via : Initial Calibration



TIC: VU034983.D

(67) 1,3,5-Trichlorobenzene
12.877min (+0.010) 0.17ug/L
response 2741

Ion	Exp%	Act%
180.00	100	100
182.00	95.00	108.46
184.00	30.20	30.39
145.00	30.80	25.03