

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100820\
 Data File : VU040575.D
 Acq On : 08 Oct 2020 20:36
 Operator : SY/MD
 Sample : L4286-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Oct 09 03:48:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 09 01:55:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	136637	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	139497	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	60461	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	57967	5.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.40%
7) Chloroethane-d5	1.92	69	46267	5.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.00%
11) 1,1-Dichloroethene-d2	2.58	63	85495	3.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.20%
20) 2-Butanone-d5	4.66	46	192538	53.14	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.28%
24) Chloroform-d	5.08	84	100275	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
26) 1,2-Dichloroethane-d4	5.72	65	63473	5.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.80%
32) Benzene-d6	5.74	84	192502	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.40%
36) 1,2-Dichloropropane-d6	6.70	67	65059	5.21	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.20%
41) Toluene-d8	7.91	98	161212	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
43) trans-1,3-Dichloropropene-	8.19	79	25375	4.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.00%
46) 2-Hexanone-d5	8.64	63	137044	50.26	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.52%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	54321	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.00%
64) 1,2-Dichlorobenzene-d4	12.19	152	63017	5.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.40%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	2393	0.191 ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100820\
Data File : VU040575.D
Acq On : 08 Oct 2020 20:36
Operator : SY/MD
Sample : L4286-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Oct 09 03:48:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 09 01:55:27 2020
Response via : Initial Calibration

