


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU100820\
 Data File : VU040554.D
 Acq On : 08 Oct 2020 09:08
 Operator : SY/MD
 Sample : BFB15
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB15

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Fri Oct 09 01:55:27 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	43349	PASS
75	95	30	80	53.9	110114	PASS
95	95	100	100	100.0	204138	PASS
96	95	5	9	6.4	13134	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.5	149952	PASS
175	174	5	9	7.4	11095	PASS
176	174	95	101	95.5	143168	PASS
177	176	5	9	6.5	9361	PASS

m/z	Abundance
36.00	2101.0
37.10	11894.0
38.10	10941.0
39.10	4193.0
40.00	300.0
43.20	145.0
44.00	1988.0
45.00	1954.0
46.00	347.0
47.00	3031.0
48.00	1508.0
49.10	9831.0
50.10	44296.0
51.10	13319.0
52.00	740.0
53.00	299.0
55.00	621.0
56.00	3696.0
57.10	6075.0
58.10	263.0
60.00	2206.0
61.10	11359.0
62.10	11101.0
63.10	9017.0
64.00	1192.0
65.10	1602.0
66.10	190.0
67.10	677.0
68.10	21576.0
69.10	22256.0
70.10	1856.0
72.10	1119.0
73.10	9796.0
74.10	38040.0
75.10	109264.0
76.10	9681.0
77.10	1290.0
78.10	843.0
79.00	8943.0
80.00	2733.0
81.00	9080.0
82.00	1947.0
83.10	308.0
86.00	322.0
87.00	8712.0
88.00	8368.0
91.00	1147.0
92.10	6431.0
93.10	9230.0
94.10	25360.0
95.10	203328.0
96.10	12921.0
97.00	458.0
102.80	109.0
104.00	1088.0
105.10	288.0
105.90	986.0
106.90	162.0
109.90	189.0
110.90	257.0
112.00	103.0
112.80	253.0
115.00	265.0
116.00	898.0
117.00	1716.0
118.00	835.0
119.00	1206.0
124.00	150.0
127.90	766.0
129.00	330.0
129.90	718.0
130.90	349.0
134.90	390.0
137.00	346.0
141.00	2221.0
142.10	227.0
143.00	2893.0
144.00	206.0
145.00	744.0
146.00	344.0
148.00	516.0
149.00	155.0
149.90	265.0
153.00	132.0
153.90	144.0
155.10	408.0
157.00	317.0
159.10	149.0
161.10	261.0
171.10	328.0
172.10	2083.0
174.10	145600.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

175.10	10776.0
176.10	137344.0
177.10	9345.0
178.00	219.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

m/z	Abundance
36.00	2068.0
37.10	12468.0
38.10	10848.0
39.10	4083.0
40.00	367.0
43.10	114.0
44.00	1970.0
45.00	1957.0
46.00	389.0
47.00	2947.0
48.00	1622.0
49.10	9902.0
50.10	44912.0
51.10	13679.0
52.00	699.0
52.90	280.0
55.10	668.0
56.00	3629.0
57.10	6326.0
58.10	351.0
60.10	2298.0
61.10	11515.0
62.10	11392.0
63.10	9251.0
64.00	1259.0
65.00	1478.0
66.10	173.0
67.10	728.0
68.10	22120.0
69.10	22728.0
70.10	1800.0
72.00	1053.0
73.10	10149.0
74.10	39112.0
75.10	113952.0
76.10	9988.0
77.10	1234.0
78.10	750.0
79.00	9077.0
80.00	2806.0
81.00	9292.0
82.00	2053.0
83.00	312.0
85.90	371.0
87.00	9121.0
88.00	8359.0
91.00	985.0
92.10	6760.0
93.10	9818.0
94.10	26224.0
95.10	211968.0
96.10	13598.0
97.00	449.0
102.90	125.0
104.00	1043.0
105.10	396.0
105.90	1022.0
106.90	163.0
109.90	173.0
110.90	201.0
112.10	114.0
112.90	276.0
115.00	308.0
115.90	866.0
117.00	1718.0
118.00	867.0
119.00	1291.0
123.00	113.0
123.90	132.0
127.90	781.0
128.90	316.0
129.90	752.0
130.90	301.0
134.90	385.0
137.00	404.0
139.90	113.0
141.00	2389.0
142.20	276.0
143.00	2941.0
144.00	160.0
145.00	726.0
146.00	343.0
148.00	471.0
149.00	147.0
149.90	300.0
153.00	133.0
154.00	144.0
155.10	412.0
157.00	373.0
159.10	164.0
161.00	307.0
171.00	282.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

172.10	1929.0
174.10	155008.0
175.10	11509.0
176.10	148288.0
177.10	9635.0
178.00	209.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

m/z	Abundance
36.00	1960.0
37.10	11652.0
38.10	9867.0
39.10	3738.0
40.00	403.0
44.00	1755.0
45.00	1676.0
46.00	412.0
47.00	2649.0
48.00	1528.0
49.00	9086.0
50.10	40840.0
51.10	12818.0
52.00	608.0
52.90	250.0
55.10	617.0
56.00	3164.0
57.10	5921.0
58.10	344.0
60.10	2199.0
61.10	10388.0
62.10	10508.0
63.10	8439.0
64.00	1156.0
65.00	1398.0
66.00	119.0
67.00	664.0
68.10	20680.0
69.10	21312.0
70.10	1588.0
72.00	941.0
73.10	9468.0
74.10	36472.0
75.10	107128.0
76.10	9219.0
77.10	1150.0
78.00	626.0
79.00	8453.0
80.00	2596.0
81.00	8704.0
82.00	1955.0
83.00	304.0
85.90	362.0
87.00	8377.0
88.00	7572.0
90.90	820.0
92.10	6251.0
93.10	9088.0
94.10	24632.0
95.10	197120.0
96.10	12885.0
97.00	407.0
103.00	152.0
104.00	952.0
105.10	444.0
106.00	1024.0
106.90	142.0
109.90	106.0
111.10	172.0
112.10	153.0
112.90	227.0
114.90	259.0
115.90	740.0
117.00	1530.0
118.00	863.0
119.00	1223.0
122.90	101.0
123.90	102.0
125.00	107.0
128.00	722.0
128.90	315.0
129.90	708.0
130.90	291.0
134.90	366.0
136.90	441.0
139.90	102.0
141.00	2317.0
142.20	278.0
143.00	2694.0
144.00	119.0
145.00	611.0
146.00	296.0
148.00	432.0
148.90	128.0
149.90	257.0
152.90	123.0
154.00	147.0
155.10	371.0
157.00	342.0
159.00	191.0
161.00	253.0
170.90	245.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

172.10	1528.0
174.10	149248.0
175.10	11000.0
176.10	143872.0
177.10	9105.0
178.00	211.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15

Scan 2875 (10.584 min): VU040554.D
BFB15

m/z	Abundance
39.80	112.0
44.10	406.0

Instrument :
MSVOA_U
ClientSampleId :
BFB15