

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100821\
 Data File : VU045252.D
 Acq On : 08 Oct 2021 15:29
 Operator : SY/MD
 Sample : M4023-05DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AJ2DL

Quant Time: Oct 09 05:32:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 09 05:29:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	280225	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	274362	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.816	152	117357	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	99601	32.468	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	64.940%		
7) Chloroethane-d5	1.890	69	97375	40.057	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	80.120%		
11) 1,1-Dichloroethene-d2	2.562	63	150014	29.120	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	58.240%#		
21) 2-Butanone-d5	4.629	46	219245	90.403	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	90.400%		
24) Chloroform-d	5.067	84	210556	43.458	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	86.920%		
26) 1,2-Dichloroethane-d4	5.707	65	151497	46.162	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.320%		
32) Benzene-d6	5.729	84	434181	45.606	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.220%		
36) 1,2-Dichloropropane-d6	6.694	67	143688	46.288	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	92.580%		
41) Toluene-d8	7.903	98	371955	45.273	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	90.540%		
43) trans-1,3-Dichloroprop...	8.182	79	62708	45.412	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.820%		
47) 2-Hexanone-d5	8.636	63	153573	89.272	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	89.270%		
56) 1,1,2,2-Tetrachloroeth...	10.761	84	227127	46.987	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	93.980%		
66) 1,2-Dichlorobenzene-d4	12.198	152	134226	52.555	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	105.100%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.575	96	13510	6.226	ug/L #	1
19) 1,1-Dichloroethane	3.871	63	5967	1.255	ug/L	98
20) cis-1,2-Dichloroethene	4.671	96	132951	52.780	ug/L	92
30) 1,1,1-Trichloroethane	5.317	97	50968	13.420	ug/L	99
34) Trichloroethene	6.546	95	91535	37.450	ug/L	99
42) Toluene	7.973	91	5802	0.561	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100821\
Data File : VU045252.D
Acq On : 08 Oct 2021 15:29
Operator : SY/MD
Sample : M4023-05DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AJ2DL

Quant Time: Oct 09 05:32:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 09 05:29:23 2021
Response via : Initial Calibration

