

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU100920\
 Data File : VU040579.D
 Acq On : 09 Oct 2020 12:45
 Operator : SY/MD
 Sample : VU1009WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK99

Quant Time: Oct 10 04:49:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 10 04:13:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	136083	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	135893	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	59995	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	53672	4.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.20%
7) Chloroethane-d5	1.92	69	43826	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.80%
11) 1,1-Dichloroethene-d2	2.58	63	81065	3.68	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.60%
20) 2-Butanone-d5	4.66	46	185702	51.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.92%
24) Chloroform-d	5.08	84	97985	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.72	65	61537	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
32) Benzene-d6	5.75	84	186899	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
36) 1,2-Dichloropropane-d6	6.70	67	61296	5.04	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.80%
41) Toluene-d8	7.91	98	155897	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	8.19	79	24774	4.81	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.20%
46) 2-Hexanone-d5	8.64	63	130642	49.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.36%
57) 1,1,2,2-Tetrachloroethane-	10.76	84	51444	4.71	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	58863	5.29	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.80%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1644	0.132 ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100920\
Data File : VU040579.D
Acq On : 09 Oct 2020 12:45
Operator : SY/MD
Sample : VU1009WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK99

Quant Time: Oct 10 04:49:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 10 04:13:39 2020
Response via : Initial Calibration

