

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100920\
 Data File : VU040595.D
 Acq On : 09 Oct 2020 19:34
 Operator : SY/MD
 Sample : L4349-02
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFSD9

Quant Time: Oct 10 05:43:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 10 04:13:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	159297	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	160515	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	70513	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	51384	3.89	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.80%
7) Chloroethane-d5	1.92	69	43459	4.27	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.40%
11) 1,1-Dichloroethene-d2	2.58	63	76663	2.97	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.40%#
20) 2-Butanone-d5	4.65	46	197339	46.72	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.44%
24) Chloroform-d	5.08	84	99099	4.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.60%
26) 1,2-Dichloroethane-d4	5.72	65	60008	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.00%
32) Benzene-d6	5.74	84	183436	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
36) 1,2-Dichloropropane-d6	6.70	67	62463	4.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.00%
41) Toluene-d8	7.91	98	151156	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.00%
43) trans-1,3-Dichloropropene-	8.19	79	24822	4.08	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.60%
46) 2-Hexanone-d5	8.64	63	141509	45.10	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.20%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	55612	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.20%
64) 1,2-Dichlorobenzene-d4	12.19	152	62239	4.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.20%

Target Compounds

					Ovalue
3) Chloromethane	1.53	50	12948	0.959 ug/L	90
13) Acetone	2.66	43	6628	2.024 ug/L	67
25) Chloroform	5.11	83	12226	0.487 ug/L	95
38) Bromodichloromethane	7.12	83	2490	0.136 ug/L #	94
47) Tetrachloroethene	8.56	164	44861	4.704 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100920\
Data File : VU040595.D
Acq On : 09 Oct 2020 19:34
Operator : SY/MD
Sample : L4349-02
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFSD9

Quant Time: Oct 10 05:43:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 10 04:13:39 2020
Response via : Initial Calibration

