

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U100920\
 Data File : VU040611.D
 Acq On : 10 Oct 2020 02:29
 Operator : SY/MD
 Sample : L4352-08
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 ETYZ7

Quant Time: Oct 10 06:52:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 10 05:58:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	131065	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	130698	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	55734	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	46066	4.24	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.80%
7) Chloroethane-d5	1.92	69	40578	4.85	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.00%
11) 1,1-Dichloroethene-d2	2.58	63	73900	3.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.60%
20) 2-Butanone-d5	4.66	46	190772	54.89	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	109.78%
24) Chloroform-d	5.08	84	97855	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
26) 1,2-Dichloroethane-d4	5.72	65	60035	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.40%
32) Benzene-d6	5.75	84	178093	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.00%
36) 1,2-Dichloropropane-d6	6.70	67	60803	5.20	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.00%
41) Toluene-d8	7.91	98	142672	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	8.19	79	22925	4.62	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.40%
46) 2-Hexanone-d5	8.64	63	139452	54.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.18%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	55044	5.24	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	57445	5.56	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.20%

Target Compounds					Ovalue
53) m,p-Xylene	9.69	106	1676	0.099 ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU100920\
Data File : VU040611.D
Acq On : 10 Oct 2020 02:29
Operator : SY/MD
Sample : L4352-08
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ETYZ7

Quant Time: Oct 10 06:52:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 10 05:58:50 2020
Response via : Initial Calibration

