

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100922\
 Data File : VU051300.D
 Acq On : 09 Oct 2022 14:42
 Operator : JC/MD
 Sample : N4931-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 10 01:27:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 10 01:25:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	179723	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	203927	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	75719	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	67990	4.690	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	93.800%	
7) Chloroethane-d5	1.912	69	65859	5.287	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	105.800%	
11) 1,1-Dichloroethene-d2	2.565	65	28270	4.933	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	98.600%	
20) 2-Butanone-d5	4.642	46	178959	56.068	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	112.140%	
24) Chloroform-d	5.060	84	150817	5.580	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	111.600%	
26) 1,2-Dichloroethane-d4	5.700	65	79242	5.685	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	113.800%	
32) Benzene-d6	5.726	84	261019	4.653	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.000%	
36) 1,2-Dichloropropane-d6	6.690	67	96098	5.044	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.800%	
41) Toluene-d8	7.896	98	198288	3.951	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	79.000%	
43) trans-1,3-Dichloroprop...	8.179	79	28077	4.486	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.800%	
46) 2-Hexanone-d5	8.636	63	93947	44.532	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	89.060%	
56) 1,1,2,2-Tetrachloroeth...	10.755	84	78379	5.150	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.000%	
66) 1,2-Dichlorobenzene-d4	12.192	152	68037	5.170	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	103.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100922\
Data File : VU051300.D
Acq On : 09 Oct 2022 14:42
Operator : JC/MD
Sample : N4931-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 10 01:27:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 10 01:25:26 2022
Response via : Initial Calibration

