

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100922\
 Data File : VU051311.D
 Acq On : 09 Oct 2022 19:45
 Operator : JC/MD
 Sample : N5013-10
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 10 01:30:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 10 01:25:26 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	2130699	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	1931679	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	820652	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	762095	4.434	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.600%
7) Chloroethane-d5	1.909	69	750252	5.080	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.600%
11) 1,1-Dichloroethene-d2	2.565	65	301627	4.440	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.800%
20) 2-Butanone-d5	4.617	46	509133	13.455	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	26.900%#
24) Chloroform-d	5.060	84	1659662	5.180	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.700	65	954185	5.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.400%
32) Benzene-d6	5.726	84	3286075	6.184	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	123.600%
36) 1,2-Dichloropropane-d6	6.687	67	1169523	6.480	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	129.600%
41) Toluene-d8	7.896	98	2733853	5.750	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.000%
43) trans-1,3-Dichloroprop...	8.176	79	382547	6.452	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	129.000%
46) 2-Hexanone-d5	8.629	63	372423	18.637	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	37.280%#
56) 1,1,2,2-Tetrachloroeth...	10.755	84	987859	6.852	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	137.000%#
66) 1,2-Dichlorobenzene-d4	12.192	152	913835	6.407	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	128.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100922\
Data File : VU051311.D
Acq On : 09 Oct 2022 19:45
Operator : JC/MD
Sample : N5013-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 10 01:30:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 10 01:25:26 2022
Response via : Initial Calibration

