

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
 Data File : VU055707.D
 Acq On : 09 Oct 2023 17:41
 Operator : MD/SY
 Sample : 04644-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :

Quant Time: Oct 10 01:32:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 10 01:10:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	10715	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	16160	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5520	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	8770	0.521	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.000%
4) 1,2-Dichloroethane-d4	5.685	65	7181	0.528	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	6.687	67	7759	0.513	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.887	98	16963	0.481	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	96.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	6274	0.555	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	112.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.597	62	436	0.029	ug/L #	1
6) Trichloroethene	6.544	95	157	0.012	ug/L	93
9) 1,2-Dibromoethane	8.924	107	51	0.006	ug/L #	84
12) 1,2,3-Trichloropropane	10.814	75	64	0.008	ug/L #	57


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
Data File : VU055707.D
Acq On : 09 Oct 2023 17:41
Operator : MD/SY
Sample : 04644-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :

Quant Time: Oct 10 01:32:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 10 01:10:06 2023
Response via : Initial Calibration









#3
 Vinyl chloride
 Concen: 0.029 ug/L
 RT: 1.597 min Scan# 114
 Delta R.T. 0.000 min
 Lab File: VU055707.D
 Acq: 09 Oct 2023 17:41




Instrument :
 MSVOA_U
 ClientSampleId :


Tgt Ion: 62 Resp: 436
 Ion Ratio Lower Upper
 62 100
 64 95.7 22.8 42.4#



#6
 Trichloroethene
 Concen: 0.012 ug/L
 RT: 6.544 min Scan# 458
 Delta R.T. 0.000 min
 Lab File: VU055707.D
 Acq: 09 Oct 2023 17:41


Tgt Ion: 95 Resp: 157
 Ion Ratio Lower Upper
 95 100
 97 84.7 47.4 88.0
 132 95.3 67.2 124.8
 130 97.6 70.1 130.1








#9
1,2-Dibromoethane
Concen: 0.006 ug/L
RT: 8.924 min Scan# 596
Delta R.T. 0.000 min
Lab File: VU055707.D
Acq: 09 Oct 2023 17:41

Instrument :
MSVOA_U
ClientSampleId :



Tgt Ion:107 Resp: 51
Ion Ratio Lower Upper
107 100
109 96.9 66.0 122.6
188 73.8 5.4 8.0#



#12
1,2,3-Trichloropropane
Concen: 0.008 ug/L
RT: 10.814 min Scan# 705
Delta R.T. 0.000 min
Lab File: VU055707.D
Acq: 09 Oct 2023 17:41

Tgt Ion: 75 Resp: 64
Ion Ratio Lower Upper
75 100
110 56.3 30.4 45.6#
77 65.6 26.2 39.4#

