

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
 Data File : VU055693.D
 Acq On : 09 Oct 2023 10:35
 Operator : MD/SY
 Sample : VSTD0.563
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD0.5063

Quant Time: Oct 10 01:04:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 10 01:03:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	11430	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	17288	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5861	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	8925	0.353	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	6561	0.463	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	7535	0.453	ug/L	0.00
8) Toluene-d8	7.887	98	17626	0.489	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	5286	0.459	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	7285	0.307	ug/L	100
6) Trichloroethene	6.544	95	6261	0.446	ug/L	100
9) 1,2-Dibromoethane	8.924	107	4122	0.509	ug/L	100
12) 1,2,3-Trichloropropane	10.814	75	3849	0.493	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.990	75	652	0.529	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
Data File : VU055693.D
Acq On : 09 Oct 2023 10:35
Operator : MD/SY
Sample : VSTD0.563
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD0.5063

Quant Time: Oct 10 01:04:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 10 01:03:32 2023
Response via : Initial Calibration

