

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
 Data File : VU055702.D
 Acq On : 09 Oct 2023 15:40
 Operator : MD/SY
 Sample : 04744-02RE
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BH579RE

Quant Time: Oct 10 01:31:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 10 01:10:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	10510	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.419	117	17966	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5689	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	27	0.002	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	0.000%#
4) 1,2-Dichloroethane-d4	5.685	65	7078	0.530	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	6.687	67	7729	0.459	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	92.000%
8) Toluene-d8	7.903	98	121	0.003	ug/L	0.02
Spiked Amount	0.500	Range	70 - 130	Recovery	=	0.000%#
10) 1,1,2,2-Tetrachloroeth...	10.745	84	6200	0.494	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	98.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
Data File : VU055702.D
Acq On : 09 Oct 2023 15:40
Operator : MD/SY
Sample : 04744-02RE
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BH579RE

Quant Time: Oct 10 01:31:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 10 01:10:06 2023
Response via : Initial Calibration

