

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
 Data File : VU055706.D
 Acq On : 09 Oct 2023 17:17
 Operator : MD/SY
 Sample : 04744-07
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Oct 10 01:32:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 10 01:10:06 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	11130	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	16900	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.808	152	5844	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	8939	0.511	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	5.685	65	7205	0.510	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	6.687	67	8016	0.506	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.887	98	17071	0.463	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.745	84	6291	0.532	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	106.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
Data File : VU055706.D
Acq On : 09 Oct 2023 17:17
Operator : MD/SY
Sample : 04744-07
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Oct 10 01:32:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 10 01:10:06 2023
Response via : Initial Calibration

