

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
 Data File : VU055694.D
 Acq On : 09 Oct 2023 10:59
 Operator : MD/SY
 Sample : VSTD00164
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VSTD001064

Quant Time: Oct 10 01:04:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 10 01:03:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.258	114	11241	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.418	117	14169	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.825	152	4951	0.500	ug/L	0.02
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.597	65	17916	0.721	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.685	65	13446	0.964	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.687	67	14848	1.090	ug/L	0.00
8) Toluene-d8	7.887	98	30307	1.026	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.745	84	9460	1.003	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.597	62	14087	0.603	ug/L	99
6) Trichloroethene	6.544	95	11817	1.028	ug/L	97
9) 1,2-Dibromoethane	8.924	107	7144	1.076	ug/L #	99
12) 1,2,3-Trichloropropane	10.814	75	6878	1.042	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.990	75	1399	1.344	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU100923\
Data File : VU055694.D
Acq On : 09 Oct 2023 10:59
Operator : MD/SY
Sample : VSTD00164
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTD001064

Quant Time: Oct 10 01:04:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM100923.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 10 01:03:32 2023
Response via : Initial Calibration

