

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\U101018\
 Data File : VU027560.D
 Acq On : 10 Oct 2018 10:23
 Operator : MD/SY
 Sample : J5200-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B

Quant Time: Oct 11 04:02:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	140364	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	239680	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	208787	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	89826	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.31	65	109069	46.53	ug/l	0.00
Spiked Amount			Recovery	=	93.06%	
35) Dibromofluoromethane	4.89	113	78842	48.00	ug/l	0.00
Spiked Amount			Recovery	=	96.00%	
50) Toluene-d8	7.57	98	300684	48.91	ug/l	0.00
Spiked Amount			Recovery	=	97.82%	
62) 4-Bromofluorobenzene	10.31	95	97078	42.49	ug/l	0.00
Spiked Amount			Recovery	=	84.98%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	2.83	59	1372	2.780	ug/l #	72
16) Acetone	2.32	43	22819	16.480	ug/l	99
18) Methyl Acetate	2.62	43	3232	1.023	ug/l	92
20) Methylene Chloride	2.70	84	3901	2.023	ug/l	97
25) 2-Butanone	4.28	43	6210	3.289	ug/l	93
29) Tetrahydrofuran	4.66	42	530	0.445	ug/l #	56
30) Chloroform	4.68	83	1148	0.351	ug/l #	76
31) Cyclohexane	4.98	56	3135	0.599	ug/l #	1
36) 1,1-Dichloropropene	4.99	75	12002	4.612	ug/l #	50
37) Ethyl Acetate	4.28	43	6210	1.846	ug/l #	70
38) Carbon Tetrachloride	4.99	117	14172	6.029	ug/l #	17
41) Methacrylonitrile	4.67	41	382	0.201	ug/l #	30
43) Isopropyl Acetate	5.55	43	125141	23.957	ug/l	99
48) Methyl methacrylate	6.71	41	9940	3.827	ug/l #	46
51) 4-Methyl-2-Pentanone	7.47	43	1105	0.326	ug/l #	75
56) Ethyl methacrylate	7.85	69	78	2.293	ug/l #	1
59) 2-Hexanone	8.38	43	1053	0.400	ug/l	92
76) 1,2,3-Trichloropropane	10.31	75	51051	21.924	ug/l #	37
80) 1,3,5-Trimethylbenzene	10.77	105	1841	0.347	ug/l	93
81) trans-1,4-Dichloro-2-buten	10.31	75	51051	57.379	ug/l #	100
84) 1,2,4-Trimethylbenzene	11.15	105	2963	0.546	ug/l	99
89) n-Butylbenzene	11.89	91	636	2.053	ug/l #	49
90) Hexachloroethane	12.29	117	246	0.236	ug/l #	12
95) Naphthalene	13.75	128	12271	3.407	ug/l	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU101018\
Data File : VU027560.D
Acq On : 10 Oct 2018 10:23
Operator : MD/SY
Sample : J5200-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OR-03-100818-B

Quant Time: Oct 11 04:02:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. 0.00 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :


OR-03-100818-B

Tgt Ion:168 Resp: 140364

Ion Ratio Lower Upper

168 100

99 61.3 50.8 76.2



Abundance Ion 167.90 (167.60 to 168.60): VU027314.D (-1169) (-)

Ion 98.90 (98.60 to 99.60): VU027314.D (-1169) (-)

4.99

60000



40000

20000

0

4.90 5.00 5.10

Time-->



#11

Tert butyl alcohol

Concen: 2.780 ug/l

RT: 2.83 min Scan# 515

Delta R.T. 0.00 min

Lab File: VU027560.D


Acq: 10 Oct 2018 10:23

Tgt Ion: 59 Resp: 1372

Ion Ratio Lower Upper

59 100

57 0.0 8.4 12.6#



Abundance Ion 59.00 (58.70 to 59.70): VU027314.D (-499) (-)

Ion 57.00 (56.70 to 57.70): VU027314.D (-499) (-)

2.83

800

600


400


200

0

2.80 2.85

Time-->





#16

Acetone

Concen: 16.480 ug/l

RT: 2.32 min Scan# 358

Delta R.T. 0.00 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :





OR-03-100818-B

Tgt Ion: 43 Resp: 22819

Ion Ratio Lower Upper

43 100

58 30.7 24.2 36.4



#18

Methyl Acetate

Concen: 1.023 ug/l

RT: 2.62 min Scan# 450

Delta R.T. 0.00 min

Lab File: VU027560.D




Acq: 10 Oct 2018 10:23


Tgt Ion: 43 Resp: 3232

Ion Ratio Lower Upper

43 100

74 17.0 16.5 24.7









#20
Methylene Chloride
Concen: 2.023 ug/l
RT: 2.70 min Scan# 476
Delta R.T. 0.00 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23




Instrument :
MSVOA_U
ClientSampleId :
OR-03-100818-B


Tgt Ion: 84 Resp: 3901
Ion Ratio Lower Upper
84 100
49 143.8 120.5 180.7
51 44.4 35.1 52.7
86 63.6 50.5 75.7



#25
2-Butanone
Concen: 3.289 ug/l
RT: 4.28 min Scan# 966
Delta R.T. 0.02 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23

Tgt Ion: 43 Resp: 6210
Ion Ratio Lower Upper
43 100
72 25.4 17.8 26.6





#29

Tetrahydrofuran

Concen: 0.445 ug/l

RT: 4.66 min Scan# 1083

Delta R.T. 0.02 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :

OR-03-100818-B





Tgt Ion: 42 Resp: 530

Ion Ratio Lower Upper

42 100

72 19.6 30.1 45.1#

71 0.0 27.7 41.5#



#30

Chloroform

Concen: 0.351 ug/l

RT: 4.68 min Scan# 1092

Delta R.T. 0.01 min

Lab File: VU027560.D




Acq: 10 Oct 2018 10:23


Tgt Ion: 83 Resp: 1148

Ion Ratio Lower Upper

83 100

85 45.7 51.8 77.6#









#31
Cyclohexane
Concen: 0.599 ug/l
RT: 4.98 min Scan# 1185
Delta R.T. -0.02 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23

Instrument :
MSVOA_U
ClientSampleId :
OR-03-100818-B

Tgt Ion: 56 Resp: 3135
Ion Ratio Lower Upper
56 100
69 197.2 24.1 36.1#
84 161.9 59.1 88.7#




Abundance Ion 56.00 (55.70 to 56.70): VU0
Ion 69.00 (68.70 to 69.70): VU0
Ion 84.00 (83.70 to 84.70): VU0






#33
1,2-Dichloroethane-d4
Concen: 46.534 ug/l
RT: 5.31 min Scan# 1287
Delta R.T. 0.00 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23

Tgt Ion: 65 Resp: 109069
Ion Ratio Lower Upper
65 100
67 52.1 0.0 107.8



Abundance Ion 64.90 (64.60 to 65.60): VU0
Ion 66.90 (66.60 to 67.60): VU0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.89 min Scan# 1467
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B

Tgt Ion:114 Resp: 239680
 Ion Ratio Lower Upper
 114 100
 63 23.5 0.0 50.2
 88 16.7 0.0 33.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VU0
 Ion 87.90 (87.60 to 88.60): VU0






#35
 Dibromofluoromethane
 Concen: 47.998 ug/l
 RT: 4.89 min Scan# 1155
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Tgt Ion:113 Resp: 78842
 Ion Ratio Lower Upper
 113 100
 111 102.3 83.1 124.7
 192 19.9 15.4 23.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.612 ug/l
 RT: 4.99 min Scan# 1186
 Delta R.T. -0.15 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B

Tgt Ion	Ratio	Lower	Upper
75	100		
110	7.8	15.5	46.5#
77	0.0	25.4	38.0#




Abundance Ion 74.90 (74.60 to 75.60): VU0
 Ion 109.90 (109.60 to 110.60): VU0
 Ion 76.90 (76.60 to 77.60): VU0






#37
 Ethyl Acetate
 Concen: 1.846 ug/l
 RT: 4.28 min Scan# 966
 Delta R.T. -0.10 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Tgt Ion	Ratio	Lower	Upper
43	100		
61	0.0	10.7	16.1#
70	0.0	7.4	11.2#



Abundance Ion 43.00 (42.70 to 43.70): VU0
 Ion 60.90 (60.60 to 61.60): VU0
 Ion 70.00 (69.70 to 70.70): VU0





#38

Carbon Tetrachloride

Concen: 6.029 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. -0.15 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :

OR-03-100818-B


Tgt Ion:117 Resp: 14172

Ion Ratio Lower Upper

117 100

119 4.5 74.5 111.7#




121 0.0 24.6 37.0#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.201 ug/l

RT: 4.67 min Scan# 1086

Delta R.T. 0.12 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Tgt Ion: 41 Resp: 382


Ion Ratio Lower Upper

41 100

39 5.2 41.1 61.7#

67 0.0 50.4 75.6#

52 0.0 20.3 30.5#






Abundance Ion 41.00 (40.70 to 41.70): VU0

Ion 39.00 (38.70 to 39.70): VU0

Ion 67.00 (66.70 to 67.70): VU0

Ion 52.00 (51.70 to 52.70): VU0





#43

Isopropyl Acetate

Concen: 23.957 ug/l

RT: 5.55 min Scan# 1362

Delta R.T. 0.00 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :

OR-03-100818-B


Tgt Ion: 43 Resp: 125141

Ion Ratio Lower Upper

43 100

61 19.5 15.3 22.9




87 10.7 8.6 13.0



Abundance Ion 43.00 (42.70 to 43.70): VU0

Ion 61.00 (60.70 to 61.70): VU0

Ion 87.00 (86.70 to 87.70): VU0



#48

Methyl methacrylate

Concen: 3.827 ug/l

RT: 6.71 min Scan# 1722

Delta R.T. 0.09 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23


Tgt Ion: 41 Resp: 9940

Ion Ratio Lower Upper

41 100

69 0.0 61.0 91.4#



39 49.8 39.9 59.9




Abundance Ion 41.00 (40.70 to 41.70): VU0

Ion 69.00 (68.70 to 69.70): VU0

Ion 39.00 (38.70 to 39.70): VU0









#50
Toluene-d8
Concen: 48.910 ug/l
RT: 7.57 min Scan# 1989
Delta R.T. 0.00 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23




Instrument :
MSVOA_U
ClientSampleId :
OR-03-100818-B


Tgt Ion: 98 Resp: 300684
Ion Ratio Lower Upper
98 100
100 65.0 52.1 78.1



#51
4-Methyl-2-Pentanone
Concen: 0.326 ug/l
RT: 7.47 min Scan# 1958
Delta R.T. 0.01 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23

Tgt Ion: 43 Resp: 1105
Ion Ratio Lower Upper
43 100
58 22.6 29.8 44.8#









#56
Ethyl methacrylate
Concen: 2.293 ug/l
RT: 7.85 min Scan# 2075
Delta R.T. -0.17 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23




Instrument :
MSVOA_U
ClientSampleId :
OR-03-100818-B


Tgt Ion: 69 Resp: 78
Ion Ratio Lower Upper
69 100
41 197.4 60.0 90.0#
39 0.0 29.5 44.3#



#59
2-Hexanone
Concen: 0.400 ug/l
RT: 8.38 min Scan# 2240
Delta R.T. 0.01 min
Lab File: VU027560.D
Acq: 10 Oct 2018 10:23

Tgt Ion: 43 Resp: 1053
Ion Ratio Lower Upper
43 100
58 57.8 26.1 78.3









#62
 4-Bromofluorobenzene
 Concen: 42.490 ug/l
 RT: 10.31 min Scan# 2841
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23




Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B


Tgt Ion: 95 Resp: 97078
 Ion Ratio Lower Upper
 95 100
 174 76.1 0.0 147.8
 176 73.5 0.0 140.8



#63
 Chlorobenzene-d5
 Concen: 50.000 ug/l
 RT: 9.09 min Scan# 2462
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Tgt Ion: 117 Resp: 208787
 Ion Ratio Lower Upper
 117 100
 82 61.7 49.2 73.8
 119 32.1 25.7 38.5





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 11.48 min Scan# 3207

Delta R.T. 0.00 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :

OR-03-100818-B


Tgt Ion:152 Resp: 89826

Ion Ratio Lower Upper

152 100

115 58.4 43.4 130.2




150 153.0 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#76

1,2,3-Trichloropropane

Concen: 21.924 ug/l

RT: 10.31 min Scan# 2841

Delta R.T. -0.19 min

Lab File: VU027560.D


Acq: 10 Oct 2018 10:23

Tgt Ion: 75 Resp: 51051

Ion Ratio Lower Upper



75 100


77 1.1 20.3 60.9#



Abundance Ion 74.90 (74.60 to 75.60): VU0

Ion 77.00 (76.70 to 77.70): VU0









#80
 1,3,5-Trimethylbenzene
 Concen: 0.347 ug/l
 RT: 10.77 min Scan# 2986
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23




Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B


Tgt Ion: 105 Resp: 1841
 Ion Ratio Lower Upper
 105 100
 120 43.7 24.1 72.4



#81
 trans-1,4-Dichloro-2-butene
 Concen: 57.379 ug/l
 RT: 10.31 min Scan# 2841
 Delta R.T. -0.20 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Tgt Ion: 75 Resp: 51051
 Ion Ratio Lower Upper
 75 100
 53 0.0 0.0 0.0
 89 0.0 0.0 0.0









#84
 1,2,4-Trimethylbenzene
 Concen: 0.546 ug/l
 RT: 11.15 min Scan# 3104
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23




Instrument :
 MSVOA_U
 ClientSampleId :
 OR-03-100818-B


Tgt Ion: 105 Resp: 2963
 Ion Ratio Lower Upper
 105 100
 120 43.5 22.1 66.5



#89
 n-Butylbenzene
 Concen: 2.053 ug/l
 RT: 11.89 min Scan# 3333
 Delta R.T. 0.00 min
 Lab File: VU027560.D
 Acq: 10 Oct 2018 10:23

Tgt Ion: 91 Resp: 636
 Ion Ratio Lower Upper
 91 100
 92 25.8 27.0 81.0#
 134 60.8 11.7 35.1#





#90

Hexachloroethane

Concen: 0.236 ug/l

RT: 12.29 min Scan# 3457

Delta R.T. 0.14 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23

Instrument :

MSVOA_U

ClientSampleId :


OR-03-100818-B

Tgt Ion:117 Resp: 246

Ion Ratio Lower Upper

117 100

201 0.0 37.9 113.6#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 200.70 (200.40 to 201.40): V

12.29

250

200



150

100

50

0

Time-->



#95

Naphthalene

Concen: 3.407 ug/l

RT: 13.75 min Scan# 3911

Delta R.T. 0.01 min

Lab File: VU027560.D

Acq: 10 Oct 2018 10:23


Tgt Ion:128 Resp: 12271

Ion Ratio Lower Upper

128 100

127 11.9 10.5 15.7

129 9.8 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

8000

6000

4000

2000

0

Time-->

