

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027537.D
 Acq On : 09 Oct 2018 21:32
 Operator : MD/SY
 Sample : J5303-22
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EB06-20181003

Quant Time: Oct 10 07:06:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	116075	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	193549	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	166587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	64279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	96990	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	4.89	113	66677	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
50) Toluene-d8	7.57	98	244736	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	10.31	95	72562	39.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.66%	
Target Compounds						
16) Acetone	2.32	43	5492	4.796	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027537.D
Acq On : 09 Oct 2018 21:32
Operator : MD/SY
Sample : J5303-22
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EB06-20181003

Quant Time: Oct 10 07:06:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

