

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027538.D
 Acq On : 09 Oct 2018 21:55
 Operator : MD/SY
 Sample : J5326-42
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PH-2

Quant Time: Oct 10 07:07:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	110742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	187576	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	163628	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	66327	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	96772	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
35) Dibromofluoromethane	4.89	113	66502	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
50) Toluene-d8	7.57	98	238687	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	10.31	95	73873	41.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.62%	
Target Compounds						
16) Acetone	2.32	43	12272	11.234	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027538.D
Acq On : 09 Oct 2018 21:55
Operator : MD/SY
Sample : J5326-42
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-PH-2

Quant Time: Oct 10 07:07:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

