

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027550.D
 Acq On : 10 Oct 2018 02:35
 Operator : MD/SY
 Sample : J5329-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 POND-1

Quant Time: Oct 10 07:59:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	134337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	233617	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	215040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	93544	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	106693	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
35) Dibromofluoromethane	4.89	113	75888	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	7.57	98	303219	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
62) 4-Bromofluorobenzene	10.31	95	100201	44.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.98%	
Target Compounds						
16) Acetone	2.32	43	13175	9.942	ug/l	98
18) Methyl Acetate	2.62	43	2896	0.958	ug/l #	89
25) 2-Butanone	4.28	43	5305	2.936	ug/l	98
43) Isopropyl Acetate	5.55	43	124126	24.379	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027550.D
Acq On : 10 Oct 2018 02:35
Operator : MD/SY
Sample : J5329-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
POND-1

Quant Time: Oct 10 07:59:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

