

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101018\
 Data File : VU027552.D
 Acq On : 10 Oct 2018 03:22
 Operator : MD/SY
 Sample : J5324-02
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OILY-DEBRIS

Quant Time: Oct 10 07:59:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	141390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	243146	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	213514	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	96079	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	109250	46.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.54%	
35) Dibromofluoromethane	4.89	113	79403	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
50) Toluene-d8	7.57	98	301461	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
62) 4-Bromofluorobenzene	10.31	95	100262	43.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.52%	
Target Compounds						
					Qvalue	
16) Acetone	2.32	43	32933	23.612	ug/l	98
18) Methyl Acetate	2.62	43	3218	1.011	ug/l	94
20) Methylene Chloride	2.70	84	60190	30.980	ug/l	97
25) 2-Butanone	4.28	43	11749	6.177	ug/l	98
43) Isopropyl Acetate	5.55	43	65300	12.323	ug/l	99
95) Naphthalene	13.74	128	14516	3.592	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027552.D
Acq On : 10 Oct 2018 03:22
Operator : MD/SY
Sample : J5324-02
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OILY-DEBRIS

Quant Time: Oct 10 07:59:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

