

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027562.D
 Acq On : 10 Oct 2018 12:07
 Operator : MD/SY
 Sample : J5302-07
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 TW-1

Quant Time: Oct 11 05:03:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	134096	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	225387	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	191747	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	77069	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	104592	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	4.89	113	74831	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
50) Toluene-d8	7.57	98	282670	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
62) 4-Bromofluorobenzene	10.31	95	87654	40.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.60%	
Target Compounds						
16) Acetone	2.32	43	6728	5.086	ug/l	99
52) Toluene	7.64	92	1443	0.342	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027562.D
Acq On : 10 Oct 2018 12:07
Operator : MD/SY
Sample : J5302-07
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
TW-1

Quant Time: Oct 11 05:03:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

