

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027571.D
 Acq On : 10 Oct 2018 15:36
 Operator : MD/SY
 Sample : PB113684ZHE#15
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 PB113684ZHE#15

Quant Time: Oct 11 05:21:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	162569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	269362	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	231308	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	96751	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	113913	41.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.92%	
35) Dibromofluoromethane	4.89	113	85895	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	7.57	98	333154	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
62) 4-Bromofluorobenzene	10.31	95	106089	41.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.64%	
Target Compounds						
16) Acetone	2.32	43	12208	7.612	ug/l	99
18) Methyl Acetate	2.62	43	3716	1.016	ug/l	94
43) Isopropyl Acetate	5.55	43	140710	23.969	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027571.D
Acq On : 10 Oct 2018 15:36
Operator : MD/SY
Sample : PB113684ZHE#15
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PB113684ZHE#15

Quant Time: Oct 11 05:21:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

