

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101018\
 Data File : VU027572.D
 Acq On : 10 Oct 2018 16:00
 Operator : MD/SY
 Sample : J5326-04
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PIT-1

Quant Time: Oct 11 05:23:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	163779	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	268729	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	231276	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	101492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	112128	41.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.00%	
35) Dibromofluoromethane	4.89	113	84883	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
50) Toluene-d8	7.57	98	326806	47.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.82%	
62) 4-Bromofluorobenzene	10.31	95	107271	41.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.76%	
Target Compounds						
16) Acetone	2.32	43	10248	6.343	ug/l	99
18) Methyl Acetate	2.62	43	3667	0.995	ug/l	96
20) Methylene Chloride	2.70	84	446877	198.569	ug/l	93
43) Isopropyl Acetate	5.55	43	129256	22.070	ug/l	99
84) 1,2,4-Trimethylbenzene	11.15	105	5774	0.942	ug/l	96
95) Naphthalene	13.74	128	249452	33.091	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027572.D
Acq On : 10 Oct 2018 16:00
Operator : MD/SY
Sample : J5326-04
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-1

Quant Time: Oct 11 05:23:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

