

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027579.D
 Acq On : 10 Oct 2018 18:43
 Operator : MD/SY
 Sample : J5326-32
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PIT-8

Quant Time: Oct 11 05:42:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	156201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	265404	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	227493	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	96899	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	110935	42.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.06%	
35) Dibromofluoromethane	4.89	113	85085	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
50) Toluene-d8	7.57	98	326222	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	10.31	95	106358	42.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.08%	
Target Compounds						
					Qvalue	
16) Acetone	2.32	43	9350	6.068	ug/l	94
18) Methyl Acetate	2.63	43	3939	1.121	ug/l #	90
20) Methylene Chloride	2.70	84	86142	40.134	ug/l	93
43) Isopropyl Acetate	5.55	43	126946	21.947	ug/l	99
95) Naphthalene	13.75	128	39538	6.878	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027579.D
Acq On : 10 Oct 2018 18:43
Operator : MD/SY
Sample : J5326-32
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-8

Quant Time: Oct 11 05:42:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration

