

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101018\
 Data File : VU027573.D
 Acq On : 10 Oct 2018 16:23
 Operator : MD/SY
 Sample : J5326-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PIT-2

Quant Time: Oct 11 05:24:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 02:04:34 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	148393	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	250743	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	220236	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	94526	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	107455	43.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.74%	
35) Dibromofluoromethane	4.88	113	80194	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
50) Toluene-d8	7.57	98	312734	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	10.31	95	101431	42.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.88%	
Target Compounds						
16) Acetone	2.32	43	8643	5.904	ug/l	92
18) Methyl Acetate	2.62	43	3448	1.033	ug/l	91
20) Methylene Chloride	2.70	84	916523	449.481	ug/l	94
43) Isopropyl Acetate	5.55	43	122598	22.435	ug/l	99
95) Naphthalene	13.75	128	28150	5.468	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101018\
Data File : VU027573.D
Acq On : 10 Oct 2018 16:23
Operator : MD/SY
Sample : J5326-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-2

Quant Time: Oct 11 05:24:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U100118W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 02:04:34 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.99 min Scan# 1186

Delta R.T. 0.00 min

Lab File: VU027573.D

Acq: 10 Oct 2018 16:23

Instrument :

MSVOA_U

ClientSampleId :


FL-PIT-2

Tot Ion:168 Resp: 148393

Ion Ratio Lower Upper




168 100

99 58.5 50.8 76.2



Abundance Ion 167.90 (167.60 to 168.60): VU027314.D (-1169) (-)

Ion 98.90 (98.60 to 99.60): VU027314.D (-1169) (-)



#16

Acetone

Concen: 5.904 ug/l

RT: 2.32 min Scan# 358

Delta R.T. 0.00 min

Lab File: VU027573.D


Acq: 10 Oct 2018 16:23

Tgt Ion: 43 Resp: 8643

Ion Ratio Lower Upper



43 100


58 34.6 24.2 36.4



Abundance Ion 43.00 (42.70 to 43.70): VU027314.D (-335) (-)

Ion 58.00 (57.70 to 58.70): VU027314.D (-335) (-)









#18
Methyl Acetate
Concen: 1.033 ug/l
RT: 2.62 min Scan# 450
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23




Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-2


Tot Ion: 43 Resp: 3448
Ion Ratio Lower Upper
43 100
74 16.6 16.5 24.7



#20
Methylene Chloride
Concen: 449.481 ug/l
RT: 2.70 min Scan# 476
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23

Tgt Ion: 84 Resp: 916523
Ion Ratio Lower Upper
84 100
49 138.5 120.5 180.7
51 42.6 35.1 52.7
86 63.7 50.5 75.7









#33
 1,2-Dichloroethane-d4
 Concen: 43.365 ug/l
 RT: 5.31 min Scan# 1287
 Delta R.T. 0.00 min
 Lab File: VU027573.D
 Acq: 10 Oct 2018 16:23




Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PIT-2


Tot Ion: 65 Resp: 107455
 Ion Ratio Lower Upper
 65 100
 67 54.1 0.0 107.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 5.89 min Scan# 1467
 Delta R.T. 0.00 min
 Lab File: VU027573.D
 Acq: 10 Oct 2018 16:23

Tgt Ion: 114 Resp: 250743
 Ion Ratio Lower Upper
 114 100
 63 23.1 0.0 50.2
 88 17.4 0.0 33.8









#35
 Dibromofluoromethane
 Concen: 46.667 ug/l
 RT: 4.88 min Scan# 1154
 Delta R.T. 0.00 min
 Lab File: VU027573.D
 Acq: 10 Oct 2018 16:23

Instrument :
 MSVOA_U
 ClientSampleId :
 FL-PIT-2

Tot Ion:113 Resp: 80194
 Ion Ratio Lower Upper
 113 100
 111 102.0 83.1 124.7
 192 20.6 15.4 23.0




Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V






#43
 Isopropyl Acetate
 Concen: 22.435 ug/l
 RT: 5.55 min Scan# 1362
 Delta R.T. 0.00 min
 Lab File: VU027573.D
 Acq: 10 Oct 2018 16:23

Tgt Ion: 43 Resp: 122598
 Ion Ratio Lower Upper
 43 100
 61 19.8 15.3 22.9
 87 11.0 8.6 13.0



Abundance Ion 43.00 (42.70 to 43.70): VU0
 Ion 61.00 (60.70 to 61.70): VU0
 Ion 87.00 (86.70 to 87.70): VU0









#50
Toluene-d8
Concen: 48.625 ug/l
RT: 7.57 min Scan# 1988
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23




Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-2


Tot Ion: 98 Resp: 312734
Ion Ratio Lower Upper
98 100
100 65.1 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 42.436 ug/l
RT: 10.31 min Scan# 2841
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23

Tgt Ion: 95 Resp: 101431
Ion Ratio Lower Upper
95 100
174 79.6 0.0 147.8
176 76.2 0.0 140.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 9.09 min Scan# 2463
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23

Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-2

Tot Ion	Ratio	Lower	Upper
117	100		
82	59.8	49.2	73.8
119	32.6	25.7	38.5




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VU0
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 11.48 min Scan# 3207
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23

Tgt Ion	Ratio	Lower	Upper
152	100		
115	58.6	43.4	130.2
150	156.0	0.0	349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#95
Naphthalene
Concen: 5.468 ug/l
RT: 13.75 min Scan# 3910
Delta R.T. 0.00 min
Lab File: VU027573.D
Acq: 10 Oct 2018 16:23

Instrument :
MSVOA_U
ClientSampleId :
FL-PIT-2

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.8	10.5	15.7
129	11.1	8.6	12.8

