

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101019\
 Data File : VU035030.D
 Acq On : 10 Oct 2019 00:09
 Operator : JC/SP
 Sample : K5265-08
 Misc : 6.36g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBAR7

Quant Time: Oct 10 07:09:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 10 04:46:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	433247	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	430811	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	175023	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	124509	35.75	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.50%
7) Chloroethane-d5	1.63	69	91549	34.88	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	69.76%#
11) 1,1-Dichloroethene-d2	2.23	63	179623	29.20	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.40%#
21) 2-Butanone-d5	4.17	46	205048	79.94	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.94%
24) Chloroform-d	4.62	84	217402	38.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	76.80%
26) 1,2-Dichloroethane-d4	5.28	65	160694	41.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.60%
32) Benzene-d6	5.31	84	496263	40.96	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.92%
36) 1,2-Dichloropropane-d6	6.31	67	165862	41.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.96%
41) Toluene-d8	7.54	98	448207	39.44	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.88%#
43) trans-1,3-Dichloropropene-	7.83	79	63732	35.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.16%
47) 2-Hexanone-d5	8.31	63	152537	80.44	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.44%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	234635	41.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.62%
64) 1,2-Dichlorobenzene-d4	11.84	152	150167	41.79	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.58%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101019\
Data File : VU035030.D
Acq On : 10 Oct 2019 00:09
Operator : JC/SP
Sample : K5265-08
Misc : 6.36g/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBAR7

Quant Time: Oct 10 07:09:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 10 04:46:10 2019
Response via : Initial Calibration

