

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101019\
 Data File : VU035021.D
 Acq On : 09 Oct 2019 20:38
 Operator : JC/SP
 Sample : K5265-20
 Misc : 7.03g/5mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 DBAT2

Quant Time: Oct 10 05:49:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100319WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Oct 10 04:46:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	406974	50.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	404368	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	163116	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	115558	35.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.66%
7) Chloroethane-d5	1.63	69	90326	36.63	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.26%
11) 1,1-Dichloroethene-d2	2.23	63	172513	29.85	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.70%#
21) 2-Butanone-d5	4.17	46	188870	78.38	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.38%
24) Chloroform-d	4.62	84	207404	39.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.00%
26) 1,2-Dichloroethane-d4	5.28	65	153379	42.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.94%
32) Benzene-d6	5.31	84	463448	40.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.50%
36) 1,2-Dichloropropane-d6	6.31	67	156033	41.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.14%
41) Toluene-d8	7.54	98	409800	38.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	76.84%#
43) trans-1,3-Dichloropropene-	7.83	79	57424	34.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.32%
47) 2-Hexanone-d5	8.31	63	144275	81.06	ug/L	0.01
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.06%
57) 1,1,2,2-Tetrachloroethane-	10.42	84	226998	42.58	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.16%
64) 1,2-Dichlorobenzene-d4	11.84	152	144031	43.01	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.02%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101019\
Data File : VU035021.D
Acq On : 09 Oct 2019 20:38
Operator : JC/SP
Sample : K5265-20
Misc : 7.03q/5mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
DBAT2

Quant Time: Oct 10 05:49:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMULM100319WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Oct 10 04:46:10 2019
Response via : Initial Calibration

