

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101022\
 Data File : VU051331.D
 Acq On : 10 Oct 2022 16:52
 Operator : JC/MD
 Sample : N5041-05
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGS21

Quant Time: Oct 11 01:02:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 11 00:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	178309	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	195960	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	76266	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	52636	3.660	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.200%
7) Chloroethane-d5	1.912	69	52754	4.269	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.400%
11) 1,1-Dichloroethene-d2	2.565	65	22046	3.877	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.600%
20) 2-Butanone-d5	4.620	46	188819	59.626	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.260%
24) Chloroform-d	5.060	84	134717	5.024	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.700	65	73752	5.333	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.600%
32) Benzene-d6	5.726	84	237038	4.397	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
36) 1,2-Dichloropropane-d6	6.690	67	88501	4.834	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.896	98	177098	3.672	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.400%
43) trans-1,3-Dichloroprop...	8.179	79	27470	4.567	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%
46) 2-Hexanone-d5	8.632	63	105201	51.894	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.780%
56) 1,1,2,2-Tetrachloroeth...	10.754	84	77459	5.296	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.000%
66) 1,2-Dichlorobenzene-d4	12.192	152	65733	4.959	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds						
17) Methyl tert-butyl Ether	3.359	73	12893	0.433	ug/L	# 93
22) cis-1,2-Dichloroethene	4.665	96	450385	29.722	ug/L	98
34) Trichloroethene	6.542	95	35780	2.192	ug/L	99
45) 1,1,2-Trichloroethane	8.401	97	6328	0.503	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101022\
Data File : VU051331.D
Acq On : 10 Oct 2022 16:52
Operator : JC/MD
Sample : N5041-05
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGS21

Quant Time: Oct 11 01:02:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 11 00:58:34 2022
Response via : Initial Calibration

