

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101022\
 Data File : VU051335.D
 Acq On : 10 Oct 2022 18:31
 Operator : JC/MD
 Sample : N5026-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0BX9

Quant Time: Oct 11 01:03:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 11 00:58:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	175229	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	202067	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	76519	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.597	65	50602	3.580	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.600%
7) Chloroethane-d5	1.909	69	53991	4.446	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.000%
11) 1,1-Dichloroethene-d2	2.565	65	21601	3.866	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.400%
20) 2-Butanone-d5	4.620	46	160855	51.689	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.380%
24) Chloroform-d	5.060	84	133964	5.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
26) 1,2-Dichloroethane-d4	5.700	65	72590	5.341	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	5.726	84	228488	4.110	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	6.690	67	88078	4.665	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.896	98	173606	3.491	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	69.800%#
43) trans-1,3-Dichloroprop...	8.179	79	25940	4.182	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.600%
46) 2-Hexanone-d5	8.632	63	88042	42.117	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.240%
56) 1,1,2,2-Tetrachloroeth...	10.755	84	74337	4.929	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.600%
66) 1,2-Dichlorobenzene-d4	12.192	152	65639	4.935	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.800%
Target Compounds						Qvalue
13) Acetone	2.626	43	10367	4.442	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101022\
Data File : VU051335.D
Acq On : 10 Oct 2022 18:31
Operator : JC/MD
Sample : N5026-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0BX9

Quant Time: Oct 11 01:03:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100722WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 11 00:58:34 2022
Response via : Initial Calibration

