

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
 Data File : VU055725.D
 Acq On : 10 Oct 2023 16:48
 Operator : MD/SY
 Sample : 04771-07ME
 Misc : 6.25g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E05J2ME

Quant Time: Oct 11 05:55:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 11 05:51:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	482104	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	493858	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	215944	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	282261	65.273	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	130.540%
7) Chloroethane-d5	1.875	69	129998	36.777	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	73.560%
11) 1,1-Dichloroethene-d2	2.534	63	391781	50.812	ug/L	-0.03
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.620%
21) 2-Butanone-d5	4.624	46	479278	149.577	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	149.580%#
24) Chloroform-d	5.055	84	545773	65.835	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	131.660%#
26) 1,2-Dichloroethane-d4	5.692	65	372590	73.019	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	146.040%#
32) Benzene-d6	5.717	84	1118283	70.772	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	141.540%#
36) 1,2-Dichloropropane-d6	6.685	67	354686	70.904	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	141.800%#
41) Toluene-d8	7.894	98	1006978	70.576	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	141.160%#
43) trans-1,3-Dichloroprop...	8.180	79	168375	70.070	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	140.140%#
47) 2-Hexanone-d5	8.656	63	352948	179.476	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	179.480%#
56) 1,1,2,2-Tetrachloroeth...	10.759	84	567267	73.673	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	147.340%#
66) 1,2-Dichlorobenzene-d4	12.193	152	348229	81.262	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	162.520%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
Data File : VU055725.D
Acq On : 10 Oct 2023 16:48
Operator : MD/SY
Sample : 04771-07ME
Misc : 6.25g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E05J2ME

Quant Time: Oct 11 05:55:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 11 05:51:51 2023
Response via : Initial Calibration

