

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
 Data File : VU055722.D
 Acq On : 10 Oct 2023 15:35
 Operator : MD/SY
 Sample : 04360-01 10X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 JLDA7

Quant Time: Oct 11 05:54:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 11 05:51:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	387029	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	386615	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.811	152	185518	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.599	65	164340	47.340	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.680%
7) Chloroethane-d5	1.888	69	139751	49.248	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	98.500%
11) 1,1-Dichloroethene-d2	2.557	63	221750	35.825	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.640%
21) 2-Butanone-d5	4.615	46	258222	100.385	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.380%
24) Chloroform-d	5.055	84	305653	45.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.860%
26) 1,2-Dichloroethane-d4	5.695	65	205460	50.156	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.320%
32) Benzene-d6	5.721	84	619811	50.106	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.220%
36) 1,2-Dichloropropane-d6	6.685	67	196025	50.057	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.120%
41) Toluene-d8	7.894	98	534176	47.824	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.640%
43) trans-1,3-Dichloroprop...	8.174	79	90016	47.852	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.700%
47) 2-Hexanone-d5	8.631	63	180402	117.182	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	117.180%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	323724	53.705	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	107.420%
66) 1,2-Dichlorobenzene-d4	12.190	152	203041	55.152	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
Data File : VU055722.D
Acq On : 10 Oct 2023 15:35
Operator : MD/SY
Sample : 04360-01 10X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
JLDA7

Quant Time: Oct 11 05:54:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 11 05:51:51 2023
Response via : Initial Calibration

