

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
 Data File : VU055728.D
 Acq On : 10 Oct 2023 18:01
 Operator : MD/SY
 Sample : 04747-06ME
 Misc : 4.40g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EX841ME

Quant Time: Oct 11 05:55:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 11 05:51:51 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	504229	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.418	117	507799	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.814	152	220038	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	198161	43.814	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.620%
7) Chloroethane-d5	1.875	69	88837	24.030	ug/L	-0.04
Spiked Amount	50.000	Range	70 - 130	Recovery	=	48.060%#
11) 1,1-Dichloroethene-d2	2.531	63	274375	34.023	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	68.040%
21) 2-Butanone-d5	4.631	46	336193	100.318	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	100.320%
24) Chloroform-d	5.055	84	380948	43.936	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.880%
26) 1,2-Dichloroethane-d4	5.692	65	257017	48.159	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.320%
32) Benzene-d6	5.714	84	764218	47.037	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.080%
36) 1,2-Dichloropropane-d6	6.685	67	246408	47.906	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.820%
41) Toluene-d8	7.894	98	697447	47.540	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.080%
43) trans-1,3-Dichloroprop...	8.181	79	114082	46.172	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.340%
47) 2-Hexanone-d5	8.660	63	236832	117.124	ug/L	0.03
Spiked Amount	100.000	Range	45 - 130	Recovery	=	117.120%
56) 1,1,2,2-Tetrachloroeth...	10.762	84	387119	48.896	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	97.800%
66) 1,2-Dichlorobenzene-d4	12.193	152	232449	53.234	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101023\
Data File : VU055728.D
Acq On : 10 Oct 2023 18:01
Operator : MD/SY
Sample : 04747-06ME
Misc : 4.40g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EX841ME

Quant Time: Oct 11 05:55:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM101023WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 11 05:51:51 2023
Response via : Initial Calibration

