

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101024\
 Data File : VU061083.D
 Acq On : 10 Oct 2024 12:27
 Operator : MD/SY
 Sample : VU1010WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK061

Quant Time: Oct 11 06:47:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 11 06:46:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	182629	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	168713	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	77017	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	50548	3.871	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.400%
7) Chloroethane-d5	1.908	69	43233	4.424	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.400%
11) 1,1-Dichloroethene-d2	2.560	65	22853	3.968	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.400%
20) 2-Butanone-d5	4.624	46	145232	46.435	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.880%
24) Chloroform-d	5.055	84	105622	4.590	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	5.695	65	57100	4.681	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	5.721	84	209290	4.705	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.682	67	69779	4.776	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.600%
41) Toluene-d8	7.891	98	185754	4.467	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
43) trans-1,3-Dichloroprop...	8.177	79	27837	4.456	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.200%
46) 2-Hexanone-d5	8.628	63	120367	46.555	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.100%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	51328	4.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	12.187	152	61345	4.981	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101024\
Data File : VU061083.D
Acq On : 10 Oct 2024 12:27
Operator : MD/SY
Sample : VU1010WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK061

Quant Time: Oct 11 06:47:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR100124WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 11 06:46:55 2024
Response via : Initial Calibration

