

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU101118\
 Data File : VU027602.D
 Acq On : 11 Oct 2018 16:30
 Operator : MD/SY
 Sample : J5164-07 2.5ppb
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 LOD-MDL-WATER-01-QT4-2018

Quant Time: Oct 12 16:34:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U101118W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 11 10:41:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	112634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	194096	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	174581	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	0.00	152	0m	50.00	ug/l	-11.48
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	92012	52.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.54%	
35) Dibromofluoromethane	4.89	113	67046	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
50) Toluene-d8	7.57	98	244406	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	10.31	95	82003	46.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.98%	
Target Compounds						
10) Methyl Iodide	2.41	142	1807	4.57	ug/l	96
11) Tert butyl alcohol	2.83	59	5469	12.87	ug/l #	93
13) Acrolein	2.20	56	4203	12.47	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU101118\
Data File : VU027602.D
Acq On : 11 Oct 2018 16:30
Operator : MD/SY
Sample : J5164-07 2.5ppb
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
LOD-MDL-WATER-01-QT4-2018

Quant Time: Oct 12 16:34:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U101118W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 11 10:41:56 2018
Response via : Initial Calibration

