

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU101118\
 Data File : VU027606.D
 Acq On : 11 Oct 2018 18:04
 Operator : MD/SY
 Sample : J5373-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 RT1499-3783-RT3026

Quant Time: Oct 12 03:07:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U101118W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 11 10:41:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.99	168	126477	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	214665	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	187854	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	73967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	101855	52.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.04%	
35) Dibromofluoromethane	4.89	113	72164	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
50) Toluene-d8	7.57	98	271080	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	10.31	95	83711	42.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.82%	
Target Compounds						
16) Acetone	2.32	43	11714	9.81	ug/l	99
20) Methylene Chloride	2.70	84	6610	3.91	ug/l	92
43) Isopropyl Acetate	5.55	43	111302	24.58	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU101118\
Data File : VU027606.D
Acq On : 11 Oct 2018 18:04
Operator : MD/SY
Sample : J5373-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
RT1499-3783-RT3026

Quant Time: Oct 12 03:07:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U101118W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 11 10:41:56 2018
Response via : Initial Calibration

