

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101119\
 Data File : VU035070.D
 Acq On : 11 Oct 2019 00:40
 Operator : JC/SP
 Sample : K5310-21 200X
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 OU2-MW-3-001-01

Quant Time: Oct 11 06:17:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 11 04:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.87	114	162816	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	170785	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	62355	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.40	65	35628	3.70	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.00%
7) Chloroethane-d5	1.68	69	37526	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.60%
11) 1,1-Dichloroethene-d2	2.26	63	55443	2.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.60%#
20) 2-Butanone-d5	4.17	46	152362	53.76	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.52%
24) Chloroform-d	4.63	84	82917	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.29	65	52121	4.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.80%
32) Benzene-d6	5.32	84	167927	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	6.31	67	62848	4.69	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.80%
41) Toluene-d8	7.55	98	144153	3.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.40%
43) trans-1,3-Dichloropropene-	7.84	79	18660	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	8.30	63	108875	47.51	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.02%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	56893	5.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.80%
64) 1,2-Dichlorobenzene-d4	11.85	152	54859	5.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.40%

Target Compounds					Ovalue
22) cis-1,2-Dichloroethene	4.21	96	4153	0.332 ug/L	98
25) Chloroform	4.65	83	13046	0.522 ug/L	96
34) Trichloroethene	6.17	95	1849	0.139 ug/L	90
47) Tetrachloroethene	8.21	164	34823	3.243 ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101119\
Data File : VU035070.D
Acq On : 11 Oct 2019 00:40
Operator : JC/SP
Sample : K5310-21 200X
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
OU2-MW-3-001-01

Quant Time: Oct 11 06:17:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 11 04:39:23 2019
Response via : Initial Calibration

