


Data Path : Z:\voasrv\HPCHEM1\MSVOA U\Data\VU101120\
 Data File : VU040633.D
 Acq On : 10 Oct 2020 13:58
 Operator : SY/MD
 Sample : BFB17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB17

Integration File: rteint.p

Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100820WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Mon Oct 12 07:06:57 2020



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2875

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	43421	PASS
75	95	30	80	55.1	114170	PASS
95	95	100	100	100.0	207253	PASS
96	95	5	9	6.3	12955	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.4	152042	PASS
175	174	5	9	7.6	11539	PASS
176	174	95	101	96.8	147136	PASS
177	176	5	9	6.3	9272	PASS

m/z	Abundance
36.00	2378.0
37.10	11913.0
38.10	10778.0
39.10	4213.0
40.00	363.0
42.90	166.0
44.00	1667.0
45.10	2289.0
45.90	306.0
47.10	3097.0
48.00	1288.0
49.10	9693.0
50.10	43536.0
51.10	13653.0
52.10	608.0
53.00	189.0
55.00	631.0
56.00	3638.0
57.10	6710.0
58.00	213.0
60.00	2460.0
61.00	11065.0
62.10	11975.0
63.10	8949.0
64.10	1190.0
65.10	1068.0
66.10	105.0
67.00	556.0
68.10	22136.0
69.10	23104.0
70.10	1954.0
72.10	1308.0
73.10	10828.0
74.10	37664.0
75.10	114600.0
76.10	9417.0
77.00	1318.0
77.90	797.0
79.00	8611.0
80.00	2507.0
81.00	8846.0
82.00	2447.0
83.10	301.0
85.90	190.0
87.10	8051.0
88.00	7493.0
91.00	835.0
92.10	6566.0
93.10	9477.0
94.10	24832.0
95.10	206656.0
96.10	13007.0
97.10	484.0
103.90	1056.0
105.00	335.0
105.90	1243.0
106.90	217.0
110.90	170.0
111.90	187.0
113.00	221.0
114.90	336.0
116.00	990.0
117.00	1667.0
117.90	785.0
119.00	1344.0
123.00	135.0
123.90	113.0
128.00	641.0
129.00	318.0
130.00	748.0
130.90	207.0
135.00	376.0
137.00	359.0
138.90	130.0
140.10	259.0
141.00	2431.0
141.90	288.0
143.00	2716.0
145.00	593.0
146.00	308.0
147.00	111.0
148.00	426.0
150.00	186.0
153.10	128.0
155.00	461.0
155.90	114.0
157.00	380.0
158.90	179.0
160.90	211.0
171.10	218.0
172.00	1650.0
174.10	146880.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

175.10	11408.0
176.10	143104.0
177.10	9113.0
178.00	224.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

m/z	Abundance
36.00	2539.0
37.00	12445.0
38.10	10815.0
39.10	4152.0
40.00	398.0
43.00	165.0
44.00	1658.0
45.10	2140.0
45.90	284.0
47.10	3292.0
48.00	1270.0
49.10	9815.0
50.10	44736.0
51.10	13860.0
52.00	669.0
53.00	171.0
55.00	649.0
56.00	3648.0
57.10	6914.0
58.10	204.0
60.00	2569.0
61.00	11287.0
62.10	11937.0
63.10	9029.0
64.10	1138.0
65.10	953.0
67.00	575.0
68.10	22808.0
69.10	24232.0
70.10	1987.0
72.00	1244.0
73.10	11633.0
74.10	38600.0
75.10	117984.0
76.10	9400.0
77.00	1372.0
77.90	809.0
79.00	9083.0
80.00	2607.0
81.00	9072.0
82.00	2373.0
83.10	310.0
87.10	8121.0
88.00	7824.0
91.00	794.0
92.10	6797.0
93.10	9637.0
94.10	26192.0
95.10	214144.0
96.10	13238.0
97.10	375.0
104.00	979.0
105.00	361.0
106.00	1359.0
106.90	237.0
110.90	197.0
111.90	162.0
113.00	220.0
114.90	332.0
116.00	1108.0
117.00	1683.0
117.90	892.0
119.00	1369.0
123.00	143.0
123.90	135.0
128.00	808.0
129.00	398.0
130.00	820.0
131.00	291.0
135.00	433.0
136.90	361.0
138.90	174.0
140.10	250.0
141.00	2549.0
141.90	262.0
143.00	2735.0
143.90	111.0
145.00	581.0
146.00	337.0
147.00	114.0
148.00	404.0
148.90	107.0
150.00	215.0
152.00	116.0
153.10	189.0
155.00	513.0
155.90	119.0
157.00	418.0
158.90	190.0
160.90	250.0
170.40	119.0
171.10	216.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

172.00	1488.0
174.10	156864.0
175.10	11968.0
176.10	152320.0
177.10	9592.0
178.00	181.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

m/z	Abundance
36.00	2362.0
37.00	11747.0
38.10	10047.0
39.10	3820.0
40.00	387.0
43.00	153.0
44.10	1641.0
45.00	1836.0
46.00	218.0
47.10	2999.0
48.00	1159.0
49.00	9049.0
50.10	41992.0
51.10	12565.0
52.00	648.0
53.00	152.0
55.10	593.0
56.10	3368.0
57.10	6313.0
58.10	227.0
60.00	2386.0
61.10	10482.0
62.10	10569.0
63.10	8360.0
64.00	1056.0
65.10	804.0
67.00	555.0
68.10	21320.0
69.10	22864.0
70.10	1776.0
72.00	1074.0
73.10	10988.0
74.10	35984.0
75.10	109928.0
76.10	8581.0
77.00	1244.0
77.90	677.0
79.00	8621.0
80.00	2530.0
81.00	8400.0
82.00	1920.0
83.00	319.0
87.10	7561.0
88.00	7114.0
90.90	700.0
92.10	6478.0
93.10	8891.0
94.10	25000.0
95.10	200960.0
96.10	12621.0
97.10	293.0
104.00	984.0
105.00	337.0
106.00	1312.0
106.90	220.0
110.10	101.0
110.90	237.0
111.90	109.0
113.00	183.0
114.90	275.0
116.00	1066.0
117.00	1602.0
117.90	923.0
119.00	1190.0
123.00	102.0
123.90	139.0
128.00	909.0
129.00	425.0
130.00	838.0
131.00	344.0
135.00	453.0
136.90	365.0
138.90	152.0
140.10	197.0
141.00	2364.0
141.90	224.0
143.00	2531.0
143.90	144.0
145.00	509.0
146.00	308.0
147.00	118.0
148.00	349.0
148.90	126.0
150.00	229.0
152.00	115.0
153.00	207.0
154.00	107.0
155.00	529.0
155.90	108.0
157.00	403.0
158.90	206.0
161.00	277.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

170.40	131.0
171.00	209.0
172.00	1305.0
174.10	152384.0
175.10	11241.0
176.10	145984.0
177.10	9113.0
178.10	180.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17

Scan 2875 (10.584 min): VU040633.D
BFB17

m/z	Abundance
39.90	178.0
44.00	276.0

Instrument :
MSVOA_U
ClientSampleId :
BFB17