


Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU101121\
Data File : VU045275.D
Acq On : 11 Oct 2021 09:06
Operator : SY/MD
Sample : BFB011
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB011

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM092321WMA.M
Title : VOC Analysis
Last Update : Wed Oct 13 02:48:52 2021



AutoFind: Scans 2890, 2891, 2892; Background Corrected with Scan 2879

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.6	3948	PASS
75	95	30	80	54.1	10339	PASS
95	95	100	100	100.0	19123	PASS
96	95	5	9	6.0	1150	PASS
173	174	0.00	2	1.2	158	PASS
174	95	50	120	70.8	13531	PASS
175	174	5	9	6.8	925	PASS
176	174	95	101	95.9	12973	PASS
177	176	5	9	6.1	785	PASS

Scan 2879 (10.597 min): VU045275.D\data.ms
BFB011

m/z	Abundance
39.80	110.0
44.00	231.0

Instrument :
MSVOA_U
ClientSampleId :
BFB011

m/z	Abundance
35.90	193.0
36.90	994.0
37.90	912.0
38.90	481.0
39.90	154.0
43.90	385.0
45.00	173.0
47.00	215.0
47.80	161.0
48.90	855.0
49.90	3942.0
50.90	1238.0
55.90	322.0
56.90	602.0
59.90	179.0
60.90	880.0
61.90	919.0
62.90	722.0
64.90	140.0
67.90	2014.0
68.90	2158.0
72.00	125.0
72.90	915.0
73.90	3449.0
75.00	10664.0
75.90	932.0
77.90	107.0
78.90	685.0
79.80	202.0
80.80	605.0
81.80	151.0
86.90	693.0
87.90	770.0
91.90	545.0
92.90	907.0
93.90	2470.0
95.00	18696.0
95.90	1159.0
116.70	110.0
118.80	106.0
140.80	168.0
142.80	178.0
171.90	108.0
172.80	122.0
173.90	12898.0
174.90	926.0
175.90	12425.0
176.80	758.0

Instrument :
MSVOA_U
ClientSampleId :
BFB011

m/z	Abundance
35.80	197.0
36.90	1046.0
37.90	1018.0
39.00	457.0
39.90	140.0
43.90	385.0
45.00	164.0
47.00	231.0
47.80	160.0
48.90	834.0
49.90	4041.0
50.90	1247.0
55.90	319.0
56.90	573.0
60.00	201.0
60.90	908.0
61.90	971.0
62.90	812.0
64.90	155.0
67.90	2077.0
68.90	2236.0
69.80	105.0
72.00	129.0
72.90	1018.0
73.90	3596.0
75.00	10605.0
75.90	924.0
76.90	113.0
77.90	109.0
78.80	681.0
79.90	195.0
80.80	641.0
81.80	199.0
86.90	784.0
87.90	776.0
91.90	592.0
92.90	967.0
93.90	2560.0
95.00	19720.0
95.90	1170.0
116.70	128.0
118.80	146.0
140.80	164.0
142.80	192.0
171.90	148.0
172.90	168.0
173.90	13850.0
174.90	945.0
175.90	13365.0
176.90	814.0

Instrument :
MSVOA_U
ClientSampleId :
BFB011

m/z	Abundance
35.90	188.0
36.90	1021.0
37.90	975.0
38.90	394.0
39.90	118.0
43.90	411.0
44.90	151.0
47.00	182.0
47.80	140.0
48.90	750.0
49.90	3861.0
50.90	1180.0
55.90	296.0
56.90	515.0
60.00	208.0
60.90	938.0
61.90	930.0
62.90	827.0
64.90	167.0
67.90	1962.0
68.90	2081.0
69.80	118.0
72.00	107.0
72.90	1001.0
73.90	3397.0
75.00	9749.0
75.90	872.0
78.80	628.0
79.90	153.0
80.80	606.0
81.80	203.0
86.90	815.0
87.90	740.0
91.90	553.0
92.90	965.0
94.00	2338.0
95.00	18952.0
95.90	1121.0
116.70	120.0
118.80	166.0
140.80	151.0
142.80	180.0
171.90	153.0
173.00	185.0
173.90	13846.0
174.80	905.0
175.90	13128.0
176.90	784.0

Instrument :
MSVOA_U
ClientSampleId :
BFB011