

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U101219\
 Data File : VU035083.D
 Acq On : 11 Oct 2019 12:22
 Operator : JC/SP
 Sample : K5310-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-10D-001-01

Quant Time: Oct 15 12:42:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Oct 12 01:49:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.86	114	162565	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.07	117	166169	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.47	152	69128	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	54528	5.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.60%
7) Chloroethane-d5	1.68	69	49875	5.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.20%
11) 1,1-Dichloroethene-d2	2.26	63	79659	3.93	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	4.19	46	163154	57.65	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.30%
24) Chloroform-d	4.62	84	100843	5.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	112.20%
26) 1,2-Dichloroethane-d4	5.29	65	54390	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.32	84	195848	5.03	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.60%
36) 1,2-Dichloropropane-d6	6.31	67	66736	5.12	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.40%
41) Toluene-d8	7.55	98	174080	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
43) trans-1,3-Dichloropropene-	7.84	79	20247	4.32	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.40%
46) 2-Hexanone-d5	8.30	63	113244	50.79	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.58%
57) 1,1,2,2-Tetrachloroethane-	10.41	84	54738	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.80%
64) 1,2-Dichlorobenzene-d4	11.84	152	60521	5.19	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.80%

Target Compounds					Ovalue
17) Methyl tert-butyl Ether	2.99	73	159925	5.614 ug/L	97
34) Trichloroethene	6.16	95	1908	0.148 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101219\
Data File : VU035083.D
Acq On : 11 Oct 2019 12:22
Operator : JC/SP
Sample : K5310-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-10D-001-01

Quant Time: Oct 15 12:42:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR100719WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Oct 12 01:49:52 2019
Response via : Initial Calibration

