

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU101219\  
 Data File : VU035118.D  
 Acq On : 12 Oct 2019 03:40  
 Operator : JC/SP  
 Sample : K5309-09  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 45 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 RB-2-191005-01

Quant Time: Oct 14 01:39:35 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR100719WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Mon Oct 14 01:16:14 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.86  | 114  | 174465   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.07  | 117  | 178495   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.47 | 152  | 74107    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.39   | 65    | 52517    | 5.09     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 101.80% |
| 7) Chloroethane-d5             | 1.68   | 69    | 48994    | 5.09     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 101.80% |
| 11) 1,1-Dichloroethene-d2      | 2.26   | 63    | 76920    | 3.54     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 70.80%  |
| 20) 2-Butanone-d5              | 4.19   | 46    | 169167   | 55.70    | ug/L | 0.02    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 111.40% |
| 24) Chloroform-d               | 4.62   | 84    | 96713    | 5.02     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 100.40% |
| 26) 1,2-Dichloroethane-d4      | 5.29   | 65    | 56598    | 4.90     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 98.00%  |
| 32) Benzene-d6                 | 5.32   | 84    | 194631   | 4.65     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 93.00%  |
| 36) 1,2-Dichloropropane-d6     | 6.31   | 67    | 66439    | 4.75     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 95.00%  |
| 41) Toluene-d8                 | 7.55   | 98    | 168700   | 4.17     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 83.40%  |
| 43) trans-1,3-Dichloropropene- | 7.84   | 79    | 19697    | 3.91     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 78.20%  |
| 46) 2-Hexanone-d5              | 8.30   | 63    | 118453   | 49.46    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 98.92%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.41  | 84    | 57621    | 4.89     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 97.80%  |
| 64) 1,2-Dichlorobenzene-d4     | 11.84  | 152   | 62310    | 4.98     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 99.60%  |

## Target Compounds

|                |      |     |       |       | Ovalue  |
|----------------|------|-----|-------|-------|---------|
| 13) Acetone    | 2.33 | 43  | 18180 | 7.309 | ug/L 93 |
| 42) Toluene    | 7.62 | 91  | 27327 | 0.487 | ug/L 99 |
| 53) m,p-Xylene | 9.36 | 106 | 2912  | 0.127 | ug/L 77 |
| 55) Styrene    | 9.78 | 104 | 4751  | 0.131 | ug/L 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU101219\  
Data File : VU035118.D  
Acq On : 12 Oct 2019 03:40  
Operator : JC/SP  
Sample : K5309-09  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 45 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
RB-2-191005-01

Quant Time: Oct 14 01:39:35 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR100719WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Mon Oct 14 01:16:14 2019  
Response via : Initial Calibration

